

**VERIFICATION OF CALIBRATION REPORT***of DataMaster cdm Breath Test Instrument**State of Alaska**Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130182</u> ✓	
Name	<u>Douglas S. Henie</u>	ID#	<u>3554</u> ✓
Date		<u>10-29-10</u> ✓	
A	Agency	<u>APD</u>	
	Phone #	<u>286-8500</u>	
	Instrument Location	<u>Anchorage TAIL</u>	
B	Alco S/N	Target Value	High Pressure
	<u>2 Q</u> ✓	<u>.080</u> ✓	<u>1050</u>
	Alco Test Values		
	<u>.079</u> ✓	<u>.080</u> ✓	
	1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco	
	Signature <u>Douglas S. Henie</u>		
	(OVER)		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
 Nita J. Bolz  
 Scientific Director  
 State Breath Alcohol Program

Subscribed and sworn before me this 15<sup>th</sup> day of DEC, 2010.

Daniel Nicholas (Notary Seal Stamp)  
 Carolyn M. Notari DANIEL NICHOLAS  
 Notary Public, State of Alaska  
 Commission Expires with Office

BT9 06/03



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130182

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130182

OCTOBER 29, 2010

OPERATOR'S NAME:

HENIE DOUGLAS S

OPERATOR'S NUMBER: 3554

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 11111111

DEPT/AGENCY: ANCI

CASE/REPORT: 10-00000

TEST TYPE: T

ALCO TARGET VALUE: .080

ALCO S/N: 2Q

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.33 in

ALCO TARGET .078

BLANK TEST .000

INTERNAL STANDARD VERIFIED

ALCO TV 29.33 in .079

BLANK TEST .000

SUBJECT SAMPLE .000

BLANK TEST .000

ALCO TV 29.33 in .080

BLANK TEST .000

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130182

OCTOBER 29, 2010

TIME 00:29

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.33 in

FLOW DETECTOR: OKAY

00:33 PUMP

00:34 HIGH SPEED: OKAY

00:34 DETECTOR: OKAY

00:35 FILTERS: OKAY

00:36 QUARTZ STANDARD: OKAY

00:37 CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGH  
IJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~  
pqrstuvwxyz{|}~  
pqrstuvwxyz{|}~